

杭州奥泰生物技术股份有限公司 Hangzhou AllTest Biotech Co.,Ltd	文件号 Document No.: ZTC-QC-005-R-003
传染病、心肌、肿瘤类 COA The Infectious Disease、Cardiology、Tumor COA	生效日期 Effective Date: 2018 年 07 月 02 日

## Certificate of Analysis

**Product Name: Strep A Rapid Test Cassette (Throat Swab)**

**Catalog No.: IST-502**

**Batch No.: STA24030020**

**Quantity:5000PCS**

**Expiry Date:2026-02**

**Date of Sampling:2024-03-12**

**Date of Analysis:2024-03-12**

**Other information:**

**Extraction reagent 1 Lot: 24035153, EXP: 2026-02**

**Extraction reagent 2 Lot: 24035154, EXP: 2026-02**

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
Functional Performance	1E+07org/ml	Positive	100% Positive	Pass
	2.5E+07org/ml	Positive	100% Positive	Pass
	Positive Control	Positive	100% Positive	Pass
	0.5%BSA-PBS	Negative	100% Negative	Pass
	Negative Control	Negative	100% Negative	Pass

<b>Others</b>	/
---------------	---

<b>Final QC Conclusion:</b>	This batch of product met the QC Criteria.
-----------------------------	--

杭州奥泰生物技术股份有限公司 Hangzhou AllTest Biotech Co.,Ltd	文件号 Document No.: ZTC-QC-005-R-003
传染病、心肌、肿瘤类 COA The Infectious Disease、Cardiology、Tumor COA	生效日期 Effective Date: 2018 年 07 月 02 日

QC supervisor: Freeman.zheng

Date: 2024.03.12



CONTROL