



GIMA

PROFESSIONAL MEDICAL PRODUCTS

BARELLA CARICAMENTO AUTOMATICO AUTOMATIC LOADING STRETCHER BRANCARD À CHARGEMENT AUTOMATIQUE AUTOMATISCH VERLADENDE KRANKENTRAGE CAMILLA DE CARGA AUTOMÁTICA MACA COM SISTEMA DE CARREGAMENTO AUTOMÁTICO ΦΟΡΕΙΟ ΑΥΤΟΜΑΤΗΣ Φ'ΟΡΤΩΣΗΣ

نقالة ذاتية التحميل

È necessario segnalare qualsiasi incidente grave verificatosi in relazione al dispositivo medico da noi fornito al fabbricante e all'autorità competente dello Stato membro in cui si ha sede.

All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.

Il est nécessaire de signaler tout accident grave survenu et lié au dispositif médical que nous avons livré au fabricant et à l'autorité compétente de l'état membre où on a le siège social.

Jeder schwere Unfall im Zusammenhang mit dem von uns gelieferten medizinischen Gerät muss unbedingt dem Hersteller und der zuständigen Behörde des Mitgliedsstaats, in dem das Gerät verwendet wird, gemeldet werden.

Es necesario informar al fabricante y a la autoridad competente del Estado miembro en el que se encuentra la sede sobre cualquier incidente grave que haya ocurrido en relación con el producto sanitario que le hemos suministrado.

É necessário notificar ao fabricante e às autoridades competentes do Estado-membro onde ele está sediado qualquer acidente grave verificado em relação ao dispositivo médico fornecido por nós.

Σε περίπτωση που διαπιστώσετε οποιοδήποτε σοβαρό περιστατικό σε σχέση με την ιατρική συσκευή που σας παρέχουμε θα πρέπει να το αναφέρετε στον κατασκευαστή και στην αρμόδια αρχή του κράτους μέλους στο οποίο βρίσκεστε.

يجب الإبلاغ فوراً عن أي حادث خطير وقع فيما يتعلق بالجهاز الطبي الذي زدنا به إلى الجهة الصناعية والسلطة المختصة في الدولة التي يقع فيها



Zhangjiagang Tengda Machinery Manufacturer Co., Ltd

No.999 Hongqi Road, Changyinsha
Modern Agricultural Demonstration Area,
Zhangjiagang City, Jiangsu, 215623, China
Made In China

GIMA 34055

REF

TD01013A



SUNGO Europe B.V.

Fascinatio Boulevard 522, Unit 1.7, 2909VA Capelle
aan den IJssel, The Netherlands



Gima S.p.A.

Via Marconi, 1 - 20060 Gessate (MI) Italy
gima@gimaitaly.com - export@gimaitaly.com
www.gimaitaly.com



Technical parameters

- Product Size: (L x W x H)
High position 193 x 60 x 88 cm
Low position 193 x 60 x 21 cm
- Maximum Angle of the Back 90°
- Carton Size 1pc 195 x 61 x 26 cm
- N.W. 40kg , G.W. 42kg
- Load Bearing 159kg

II. How to use

- The stretcher surface is made of soft sponge mattress, which make patient feel comfortable. One medical staff can push the trolley on to the ambulance or send to ward.
- When need low position, pull out the safety handle & then press it, the stretcher's height will automatically lower down.
- When destination arrived, pulling out the safety handle, the stretcher's legs will unfold automatically.
- There are the return elastic lock device at both side of the stretcher & the leg parts. Lock tightly when turn up for patient safety, turn down for patients out.
- Pressing the button under the stretcher, the head part will be raised or lowered.

III. Applied for

Approved by clinic practice, this item is suitable for ambulance, hospital battlefield gym, etc. to carry the wounded & patients. It also can be used in general operation.

IV. Cautions

- When carrying patients with this stretcher, be sure lock tightly the fender locking device & fasten the safety belt for safety sake.
- In general operation, be careful not to out torn the stretcher surface.

V. Maintenance

- Keep clean as routine (incl. sterilization).
- Often check if any parts loose or not.

VI. Storage & transportation

- Store this product in the place of damp proof & non corrosion.
- Common transporting vehicle can carry this product.
- Upside down is prohibited during transportation & storage. Stack the goods as per the instructions printed on the package.

	Product code		Manufacturer		Consult instructions for use
	Lot number		Medical Device compliant with Regulation (EU) 2017/745		Date of manufacture
	Keep in a cool, dry place		Caution: read instructions (warnings) carefully		Authorized representative in the European Union
	Keep away from sunlight		Medical Device		Unique identifier
	Imported by				

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.