

DECLARATION OF CONFORMITY

PVS SpA located in Leonardo da Vinci, 18 street , 20060 Cassina De' Pecchi (Mi), Italy, as assembler of the following kit :

Description	MULTISAN KIT ALL.1 DM 388
REF	CPS200
Basic -UDI-DI	80340280114D SRN - IT - PR - 000050312


Composition :

- 1 Copy DM 388 of 15/07/03,
- 5 Pairs of sterile gloves, MD class Is
- 1 Mask with visor , MD class I
- 3 Bottles of sterile saline solution 500 ml CE, MD class IIa
- 2 Bottles of disinfectant 500 ml iodopovidone 10% iodine PMC,
- 10 bags of sterile gauze compressed 10x10 cm, MD class Is
- 2 Bags of sterile compressed gauze 18x40 cm, MD class Is
- 2 Sterile sheets cm 40x60 DIN 13152-BR, MD class Is
- 2 Sterile tweezers MD class Is
- 1 Pack of cotton wool, MD class I
- 1 Packs of 10 assorted patches, MD class I
- 2 Spools adhesive patch m 5x2,5 cm, MD class I
- 1 Pair of scissors cutting bandages 14.5 cm DIN MD class I
- 3 Tourniquets MD class I
- 2 Instant Ice Pack MD class IIa
- 2 Bags for medical waste
- 1 Clinical thermometer MD class IIa
- 1 Personal Sphygmomanometer with phonendoscope. MD class Im

In compliance with Article 22 of REG. UE 2017/745, declares that :

- a) the mutual compatibility of the devices has been verified;
- b) is responsible for assembly, packaging and has provided users with the relevant information containing the relevant manufacturer's instructions
- (c) the activity of combining the devices into procedural kits has been subjected to adequate methods of internal control, verification and validation.
- d) the system incorporates both CE medical devices and non-medical devices.

LEGAL REPRESENTATIVE



IRENE PEREGO

Cassina De Pecchi, 30/09/2025