

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

 MANUFACTURER:	CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
MEDICAL DEVICE:	Pocket Fetal Doppler probe , DCD8E30
CLASSIFICATION - ANNEX IX:	Class II a, Rule 10
CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4	
WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.	
STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.	
NOTIFIED BODY:	TÜV SÜD PRODUCTS SERVICE GMBH RIDLERSTR65, D-80339 MÜNCHEN, GERMANY
IDENTIFICATION NUMBER:	 0123
(EC) CERTIFICATE(S):	<u>G1 050972 0050 Rev.04</u>
EC REP	
EUROPEAN REPRESENTATIVE:	Prolinx GmbH Brehmstr.56, 40239, Duesseldorf, Germany

START OF CE-MARKING: 2009-07-23 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION: QINHUANGDAO, 2024-09-13

SIGNATURE:



President

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

Appendix: list of (harmonised - EN) standards

No.	Serial Number	Title and Description
1	EN 60601-1:2006 (IEC 60601-1:2005)	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
2	EN 60601-1-2: 2007 (IEC60601-1-2:2007)	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN 60601-1-8:2007 (IEC 60601-1-8:2006)	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance -Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
4	EN 60601-1-6:2010 (IEC 60601-1-6:2010)	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability
5	EN 60601-2-37:2008 (IEC 60601-2-37:2007)	Medical electrical equipment - Part 2-37: Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment
6	EN 62366:2008 (IEC 62366:2007)	Medical devices - Application of usability engineering to medical devices
7	EN 62304:2006 (IEC 62304:2006)	Medical device software - Software life-cycle processes