


**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

MANUFACTURER:	CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
MEDICAL DEVICE:	SPIROMETER SP80B
CLASSIFICATION - ANNEX IX:	Class II a, Rule 10
CONFORMITY ASSESSMENT ROUTE:	Annex II excluding chapter 4
<p>WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.</p>	
<p>STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.</p>	
NOTIFIED BODY:	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY
IDENTIFICATION NUMBER:	CE 0123
(EC) CERTIFICATE(S):	<u>G1 050972 0050 Rev.04</u>
EUROPEAN REPRESENTATIVE:	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING: 2019-11-07 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:	QINHUANGDAO, 2020-06-18
SIGNATURE:	 _____ President

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	IEC 60601-1:2005/A1:2012	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
2	IEC 60601-1-2:2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests
3	EN 60601-1-11:2015	Medical electrical equipment -- Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
4	EN 62304:2006	Medical device software-Software life-cycle processes
5	ISO 23747:2015	Anaesthetic and respiratory equipment -- Peak expiratory flow meters for the assessment of pulmonary function in spontaneously breathing humans
6	ISO 26782:2009	Anaesthetic and respiratory equipment --Spirometers intended for the measurement of time forced expired volumes in humans