

## **Certificate of Conformity**

NO.: ES131031186E-2

The submitted sample of the following equipment has been tested for CE marking according to the following European Directives: Medical Devices Directive 93/42/EEC.

**Applicant** 

: SHENZHEN CREATIVE INDUSTRY CO.,LTD.

Address

: 2/F, Block 3, Nanyou Tian'an Industry Town, Shenzhen,

Guangdong, 518054 CHINA

Manufacturer

: SHENZHEN CREATIVE INDUSTRY CO.,LTD.

Address

: 2/F, Block 3, Nanyou Tian'an Industry Town, Shenzhen,

Guangdong, 518054 CHINA

Trademark

: N/A

**EUT** 

: Capnograph and Oximeter

M/N

: PC-900B

Test Report:

: ES131031186E-2

Test Standards

: CISPR 11:2009 +A1:2010

IEC 60601-1-2:2007 EN 61000-3-2:2014 EN 61000-3-3:2013

CE

(Manager)
July 26, 2017 FICH

The certificate is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. logo.