



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- Regulation 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, and
- Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment, as amended by Commission Delegated Directive (EU) 2015/863 of 31 March 2015 (RoHS3).

Document Number 80016309

Version R

Product Name

Power Handles & Chargers

Manufacturer's Name and
Business Address

Welch Allyn, Inc.
4341 State Street Road
Skaneateles Falls, NY 13153
USA

SRN: US-MF-000013394

Declaration of Conformity
Validity

ISO 13485 #314505 MP2016 Expiry Date: 2022-12-08



Welch Allyn Limited,
Navan Business Park, Dublin Road,
Navan, Co. Meath, C15 AW22
Ireland

SRN: IE-AR-000000768

Object of the declaration



Intended Purpose

The handles are intended to power various accessory heads including otoscopes, ophthalmoscopes, retinoscopes, strabismoscopes, episcopes, illuminators and transilluminators.

Medical Device Conformity
Assessment Route Annex

Annex II and Annex III

Medical Device Classification

Class I

Medical Device Classification
Rule

Rules 1, 13

Standards

Refer to Appendix A



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

REF #		901087: INSTRUMENT HANDLE			
		71000	71001- B	71001-C	71010
		71020	71021-C	71062	71064
		71066	71670	71902	71904
		71906	71907	71910	71911
		72800	72801	72830	71000- A
		71000-B	71000-C	71020-A	71020-B
		71020-C	71930	71900-USB	
GMDN Code and Term		34158 Secondary battery			
UMDNS Code and Term		18557 Power Supplies			
Basic UDI-DI		0732094GMN901087FL			



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Accessories

Object of the declaration



Intended Purpose 719 USB Charger, Universal Desk Charger
The chargers are intended to charge Welch Allyn rechargeable 2.5 V and 3.5 V handles.

Medical Device Conformity Assessment Route Annex Annex II and Annex III

Medical Device Classification Class I

Medical Device Classification Rule Rule 13

Standards Refer to Appendix A

REF

#

901001: ACCESSORY, EYE, EAR, NOSE & THROAT
901087: INSTRUMENT HANDLE

71142	71943
71942	71960-POD

71712	71714	71716
71732	71734	71736
71062-C	71064-C	71140
71144	71146	71955

GMDN Code and Term 17115 Non-invasive Device Battery Charger

UMDNS Code and Term 18557 Power Supplies

Basic UDI-DI 0732094GMN901001EG
0732094GMN901122EV



Hillrom™

DECLARATION OF CONFORMITY

(in accordance with ISO/IEC 17050-1)

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Approval

A handwritten signature in blue ink, appearing to read 'JK', written over a horizontal line.

Joshua Kim, Sr. Manager Global Regulatory Affairs

2022-01-31

Date

Skaneateles Falls NY,
USA

Place of Issue



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Appendix A: Standards and Common Specifications

Standards Applied	Number	Version/Date of Issue	Title
Regulation 2017/745	EN 60601-1	2014	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
	EN ISO 13485	2016	Medical Devices - Quality Management Systems - Requirements for Regulatory Purposes
	EN ISO 10993-1	2020	Biological evaluation of medical devices_ - Part_1: Evaluation and testing within a risk management process
	EN ISO 10993-5	2009	Biological evaluation of medical devices_ - Part_5: Tests for in vitro cytotoxicity
	EN ISO 10993-10	2013	Biological evaluation of medical devices_ - Part_10: Tests for irritation and skin sensitization
	EN 60601-1-2	2015	Medical electrical equipment_ - Part_1-2: General requirements for basic safety and essential performance_ - Collateral standard: Electromagnetic compatibility_ - Requirements and tests
	EN 60601-1-6	2015	Medical electrical equipment_ - General requirements for basic safety and essential performance_ - Collateral Standard: Usability
	EN 62366-1	2015	Medical devices_ - Application of usability engineering to medical devices
	EN 62304	2015	Medical device software - Software life-cycle processes
	EN ISO 15223-1	2016	Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General Requirements
Directive 2011/65/EU + (EU) 2015/863	EN IEC 63000	2018	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances