



DECLARATION OF CONFORMITY

PVS SpA located in Leonardo da Vinci, 18 street , 20051 Cassina De' Pecchi (Mi), Italy, as assembler of the following kit :

Description	MEDICAL BAG
REF	CPS282
Basic -UDI-DI	80340280114D SRN – IT- PR- 000050312

Composition :

- 1 Isothermic blanket Gold/Silver 160x210 cm MD class I
- 1 Disinfectant liquid ml.250
- 2 Adhesive plasters m 5 x 2,5 cm MD class I
- 2 Pairs of sterile gloves MD class Is
- 2 VENTO mouth to mouth resuscitators MD class I
- 6 Sterile gauze pads cm 18x40 MD class Is
- 2 Instant ice sprays ml 200 MD class IIa
- 3 Triangular bandages TNT cm 96x96x136 MD class I
- 3 Elastic bandages cm 6 MD class I
- 3 Elastic bandages cm 8 MD class I
- 1 Cotton wool pack MD class I
- 1 Pair of LISTER scissors cm 14,5 MD class Is
- 2 Compressive pad bandages mm 80x100 MD class I
- 2 Compressive pad bandages mm 100x120 MD class I
- 1 Pack of 20 adhesive plasters cm 2x7 MD class I
- 8 Adhesive plasters cm 10x6 MD class I
- 1 Sterile bandage TNT cm 60x40 MD class Is
- 1 Sterile bandage TNT cm 60x80 MD class Is
- 3 Sterile gauzes 10x10 cm 8 plies MD class Is
- 5 Instant ice packs MD class IIa
- 1 Hydrogen peroxide solution
- 1 Elastic net bandage MD class I
- 1 Haemostatic lacet MD class I
- 1 Sterile tweezers MD class Is
- 5 Ammonia towelettes
- 1 Sterile saline solution bottle ml 250 MD class IIa
- 1 Multilingual first aid guide

In compliance with Article 22 of REG. UE 2017/745, declares that :

- a) the mutual compatibility of the devices has been verified;
- b) is responsible for assembly, packaging and has provided users with the relevant information containing the relevant manufacturer's instructions
- (c) the activity of combining the devices into procedural kits has been subjected to adequate methods of internal control, verification and validation.
- d) the system incorporates both CE medical devices and non-medical devices.

Cassina De Pecchi, 04/10/25

LEGAL REPRESENTATIVE
IRENE PEREGO
