

SEJOY	文件名称 File Name	SARS-CoV-2&Influenza A+B Antigen Combo Rapid Test Cassette Completed Product Test Report Form 新冠&流感检测试剂盒成品报告单
	文件编号 File Number	Q4-000ICA-PZ-115A
		NO. <i>MP15A20038001</i>
		版本 Version
		A/0

功能性检查 Functional checks		阴性符合率 Negative compliance rate	Flu A 阳性符合率 Flu A positive compliance rate	Flu B 阳性符合率 Flu B positive compliance rate	COVID-19 阳性符合率 COVID-19 positive compliance rate
合格 <input checked="" type="checkbox"/> 合格 不合格 <input type="checkbox"/> 不合格 Unqualified	Flu A	T < G3	1	1	1
	Flu B	A < G3	1	1	1
	COVID-19	B > G3	1	1	1
合格 <input checked="" type="checkbox"/> 合格 不合格 <input type="checkbox"/> 不合格 Unqualified	Flu A	G3 ≤ A ≤ G5	3	3	3
	Flu B	B < G3	1	1	1
合格 <input checked="" type="checkbox"/> 合格 不合格 <input type="checkbox"/> 不合格 Unqualified	Flu A	G4 ≤ A ≤ G6	4.5	4.5	4.5
	Flu B	Reference panel H3N2	1	1	1
合格 <input checked="" type="checkbox"/> 合格 不合格 <input type="checkbox"/> 不合格 Unqualified	Flu B	Flu B Reference panel Flu B	4	4	4
	Flu A	A < G3	1	1	1
	COVID-19	T ≥ G7	7.5	7.5	7.5
合格 <input checked="" type="checkbox"/> 合格 不合格 <input type="checkbox"/> 不合格 Unqualified	Flu B	G4 ≤ A ≤ G6	4.5	4.5	4.5
	Flu A	Reference panel S2	1	1	1
	COVID-19	T ≥ G4.5	4.5	4.5	4.5
合格 <input checked="" type="checkbox"/> 合格 不合格 <input type="checkbox"/> 不合格 Unqualified	Flu B	Reference panel S1	1	1	1
	Flu A	Reference panel S3	1	1	1
	COVID-19	T ≥ G3	3.5	3.5	3.5

SEJOY	文件名称 File Name	SARS-CoV-2&Influenza A+B Antigen Combo Rapid Test 新冠&流感检测试剂盒成品报告单 Cassette Completed Product Test Report Form	版本 Version	A/0
	文件编号 File Number	Q4-000ICA-PZ-115A	NO.	MP15FA240308001

重复性 Repeatability	参考上述检测项目，判阳性企业内部参考品检测的判读结果是否一致，且显色度是否均一 Referring to the above test items, the readings of the internal reference tests of the positive companies are consistent and the colour rendering is uniform	<input checked="" type="checkbox"/> 合格 <input type="checkbox"/> 不合格
	综合判定 Comprehensive determination	<input checked="" type="checkbox"/> 合格 Qualified <input type="checkbox"/> 不合格 Unqualified
备注 Remark	如检验不合格，备注是否复检，并附上复检结果。 If the test fails, make a note of whether to retest and attach the results of the retest.	
检验员 Inspector:	日期 Date: 2024.07.08	复核员 Reviewer: 付冬梅
日期 Date: 2024.3-8		