

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

	MANUFACTURER: CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
MEDICAL DEVICE:	Temperature Probe, CGP0002
CLASSIFICATION - ANNEX IX:	Class II b, Rule 10
CONFORMITY ASSESSMENT ROUTE:	Annex II excluding chapter 4
We, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.	
STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.	
NOTIFIED BODY:	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY
IDENTIFICATION NUMBER:	 0123
(EC) CERTIFICATE(s):	<u>G1 050972 0050 Rev.04</u>
	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany
EUROPEAN REPRESENTATIVE:	

START OF CE-MARKING: 2012-04-20 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:	QINHUANGDAO, 2020-06-18
SIGNATURE:	 President